

CALIBRATION CERTIFICATE

Customer Details-

Burton Skip Hire Limited
The Cottage
Rough Hayes
Burton on Trent
Staffordshire
DE13 9PX

Cal Cert Ref-	<i>100550351-29/10/2019</i>
----------------------	-----------------------------

<i>Scale Details</i>		<i>As Found Linearity & Hysteresis</i>		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	100550351	0kg	0kg	0
Manufacturer	Avery E1205 BMS	4000kg	4000kg	0
Capacity	40000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	29/10/2019	16000kg	16000kg	0
Next Calibration Due	October 2020	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	28040kg	28040kg	0
Issue date	29/10/2019	38280kg	38280kg	0

REPEATABILITY				
LOAD	READING 1	READING 2	READING 3	READING 4
38280kg	38280kg	38280kg	38280kg	38280kg

ECCENTRIC LOADING							
POSITION	LOAD	READING	INDICATOR				
1	6240kg	6240kg	<table border="1"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>3</td> <td>4</td> </tr> </table>	1	2	3	4
1	2						
3	4						
2	6240kg	6240kg					
3	6240kg	6240kg					
4	6240kg	6240kg					

Compliance Weights Used		Calibration of above machine	
Box Set No	Cal Cert No	Tested By-	John Bushall
PM51-PM75, 47-50,1 & 2	W19001/W19002	Signature-	<i>J Bushall</i>
All test weights are calibrated and traceable to National Standards.			
Make up weights are also used FLT -4220kg			