

# CALIBRATION CERTIFICATE

**Customer Details-**


**Burton Skip Hire Limited  
The Cottage  
Rough Hayes  
Burton on Trent  
Staffordshire  
DE13 9PX**

<b>Cal Cert Ref-</b>	<i>100550351-09/11/2023</i>
----------------------	-----------------------------

<i>Scale Details</i>		<i>As Found Linearity &amp; Hysteresis</i>		
Location	Yard	LOAD	READING	DEVIATION
<b>Serial Number</b>	<b>100550351</b>	<b>0kg</b>	0kg	0
<b>Manufacturer</b>	Avery E1205 BMS	<b>4000kg</b>	4000kg	0
<b>Capacity</b>	40000kg x 20kg	<b>8000kg</b>	8000kg	0
<b>Machine Type</b>	Weighbridge	<b>12000kg</b>	12000kg	0
<b>Calibration Date</b>	09/11/2023	<b>16000kg</b>	16000kg	0
<b>Next Calibration Due</b>	<b>November 2024</b>	<b>10000kg</b>	10000kg	0
<b>Tolerance</b>	+/- 3 Divisions	<b>20000kg</b>	20000kg	0
<b>Issue date</b>	09/11/2023	<b>24660kg</b>	24660kg	0

<i>REPEATABILITY</i>				
LOAD	READING 1	READING 2	READING 3	READING 4
<b>24660kg</b>	24660kg	24660kg	24660kg	24660kg

<i>ECCENTRIC LOADING</i>									
POSITION	LOAD	READING	INDICATOR						
<b>1</b>	<b>6660kg</b>	6660kg	<table border="1"> <tr><td>1</td><td>2</td></tr> <tr><td>3</td><td>4</td></tr> <tr><td>5</td><td>6</td></tr> </table>	1	2	3	4	5	6
1	2								
3	4								
5	6								
<b>2</b>	<b>6660kg</b>	6660kg							
<b>3</b>	<b>6660kg</b>	6660kg							
<b>4</b>	<b>6660kg</b>	6660kg							
<b>5</b>	<b>6660kg</b>	6660kg							
<b>6</b>	<b>6660kg</b>	6660kg							

<i>Compliance Weights Used</i>		<i>Calibration of above machine</i>	
Box Set No	Cal Cert No	Tested By-	Jay Hatton
<b>W23017 &amp; W23018 - 02/2024</b>	<b>W23017/W23018</b>	<b>Signature-</b>	

**All test weights are calibrated and traceable to National Standards. | Make up weights are also used as required**