

CALIBRATION CERTIFICATE

Customer Details-


Burton Skip Hire Limited
The Cottage
Rough Hayes
Burton on Trent
Staffordshire
DE13 9PX

| | |
|----------------------|----------------------|
| Cal Cert Ref- | 100550351-22/11/2018 |
|----------------------|----------------------|

| <i>Scale Details</i> | | <i>As Found Linearity & Hysteresis</i> | | |
|-----------------------------|----------------------|--|---------|-----------|
| Location | Yard | LOAD | READING | DEVIATION |
| Serial Number | 100550351 | 0kg | 0kg | 0 |
| Manufacturer | Avery E1205 BMS | 4000kg | 4000kg | 0 |
| Capacity | 40000kg x 20kg | 8000kg | 8000kg | 0 |
| Machine Type | Weighbridge | 12000kg | 12000kg | 0 |
| Calibration Date | 22/11/2018 | 16000kg | 16000kg | 0 |
| Next Calibration Due | November 2019 | 20000kg | 20000kg | 0 |
| Tolerance | +/- 3 Divisions | 34140kg | 34140kg | 0 |
| Issue date | 22/11/2018 | 38360kg | 38360kg | 0 |

| <i>REPEATABILITY</i> | | | | |
|----------------------|-----------|-----------|-----------|-----------|
| LOAD | READING 1 | READING 2 | READING 3 | READING 4 |
| 38360kg | 38360kg | 38360kg | 38360kg | 38360kg |

| <i>ECCENTRIC LOADING</i> | | | | | | | |
|--------------------------|----------------|---------|---|---|---|---|---|
| POSITION | LOAD | READING | INDICATOR | | | | |
| 1 | 16000kg | 16000kg | <table border="1"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>3</td> <td>4</td> </tr> </table> | 1 | 2 | 3 | 4 |
| 1 | 2 | | | | | | |
| 3 | 4 | | | | | | |
| 2 | 16000kg | 16000kg | | | | | |
| 3 | 16000kg | 16000kg | | | | | |
| 4 | 16000kg | 16000kg | | | | | |

| Compliance Weights Used | | Calibration of above machine | |
|--|------------------|-------------------------------------|---|
| Box Set No | Cal Cert No | Tested By- | Mark Foxall |
| 01-22,47-50,1&2, PM51-PM75 | W17195,W18012-13 | Signature- |  |
| All test weights are calibrated and traceable to National Standards. | | | |
| Make up weights are also used FLT-4220kg Loading Bucket – 14140kg | | | |